

Averill Park Central School

**Attn: Aaron Heffner
146 Gettle Road St. 1
Averill Park ,NY 12018**

Printed On : 9/25/2025 Page 2 of 2

Sample ID: BG10630
Date Received: 08/21/2025
Time Received: 11:15
Date Finalized: 9/25/2025
PO Number:
Your Ref:

Customer: Averill Park Central School
Owner: Averill Park Central School
Sample Loc: Algonquin Middle School
Sample Pt: MID GAC

Collect Date: 08/21/2025
Collect Time: 08:55
Collected by: BRIAN COLLINS
Receipt Temp: 4.1 C On Ice Chilling

3RD QUARTER
PFOA
ALGONQUIN MID GAC

Water Source: Drilled Well
Chlorinated: Yes **Field Residual Chlorine:**

Potable: Yes
Grab/Comp: Grab

Qualifiers Key:

- | | | |
|---|--|-----------------------------|
| X Exceeds maximum contamination limit | R Duplication outside acceptance limits | H Hold time exceeded |
| T Temperature outside specifications | A Sample contained air bubble or headspace | B Analyte detected in blank |
| C(+/-) CCV outside acceptance limits | Z Analysis is not state-certified | G Incorrect bottle received |
| S(+/-) Lab control sample outside acceptance limits | M(+/-) Matrix spike recovery outside acceptance limits | P Sample preserved at lab |
| J Analyte detected below quantitation limit | I(+/-) IS/Surrogate outside acceptance limits | |

Legend: < Less Than, > Greater Than mg/L=PPM, ug/L=PPB If no collection time was given, 00:00 is reported

MCL = Maximum Contaminant Level referenced from New York State Subpart 5-1 of the Public Drinking Water Standards and/or National Primary/Secondary Drinking Water Standards.

Note 1: Per ELAP requirements, water analyzed for alkalinity, color, conductivity, nitrate, nitrite, sulfate, organics, UV absorbance, non-potable bacteriological analyses, BOD/CBOD, solids and phosphorus are required to be on ice to indicate the chilling process has begun. Samples must be between 0-6C and not frozen.

Comments:

PFAS: SUB* PFAS analyses were completed by NYS DOH Lab. #12058. Samples were prepared on 08/28/25.
Surrogates: Surrogate recovery for surrogate M3HFPO-DA was below acceptable limits. Isotope dilution applied. All other surrogate recoveries within acceptable limits.
PFAS FIELD BLANK:
All analytes - None Detected
Surrogates: All surrogate recoveries within acceptable limits.

Test procedures for all analyses meet NELAC requirements unless noted. If you have any questions, please call the laboratory.

Brian Collins
Lead Technical Director Environmental Laboratory
and contact person
If you have questions, please call.

Reviewed by Brian Collins
These results relate to samples as received.